*International Journal of Basic and Applied Sciences*, 2 (4) (2013) 336-339
©Science Publishing Corporation
www.sciencepubco.com/index.php/IJBAS

# Radiation dose measurement and environmental surveying during x-ray imaging and fluoroscopy at general hospital in Sulaimania-Kurdistan region-Iraq

**Kharman Akram Faraj**

*Department of physics, faculty of science and science education, school of science, university of sulaimania, kurdistan region , iraq*
**Corresponding author E-mail: kharmanakram@gmail.com*

**Abstract**

Radiation dose (rates) measurement and environmental surveying in X-ray units are essential to ensure that the radiographers and the visitors to the hospitals are not received excess X-ray. In the present study the dose rates were measured at radiographers place and in the front of the direct door of X-ray room during X-ray imaging and fluoroscopy at General Hospital in Sulaimania and visual inspection for radiographer's facilities were checked. High dose rates were recorded during X-ray and fluoroscopy and the issue of safety for the workers did not exist. The responsible of the hospital was informed about the results.

***Keywords***: *Dose rates, environmental surveying, fluoroscopy, X-ray.*

## 1 Introduction

Since the end of the 19th Century, man has learned to use radiation for many beneficial purposes. Today, many sources of radiation, such as x-ray machines, linear accelerators and radionuclides are used in clinical and research applications. Such beneficial uses may at times create potentially hazardous situations for personnel who work within the hospital [1].
X-ray is used daily in hospitals and clinics to perform diagnostic imaging procedures and radiographers are occupationally exposed at low-level dose and dose rate (external exposure) to X- rays, so if they do not apply the rules of radiation protection may be receive excess dose during their lifetimes. Any increment of exposure above the natural background levels will produce a linear increment of risk [2]. Radiation workers are predicted to have a greater percentage risk of developing detrimental effects over the general public because of their generally greater exposure [3].
It has been proved that ionizing radiation produces Reactive Oxygen Series (ROS) in biological system capable of destroying biomolecules such as DNA, lipids, proteins and carbohydrate [4). Increased frequencies of chromosome aberrations and micronuclei are well known among individuals occupationally exposed to ionizing radiation [5, 6, 7, 8, and 9] and high levels of chromosomal aberrations in the lymphocytes of medical staff occupationally exposed to X-rays were reported by Kasuba et al. [10]. Lalic et al. and Cheriyan et al. [11, 12] observed a high incidence of chromosomal aberrations in human population from southwest coast of India who were exposed to low level natural radiations. In another study high incidence of leukemia and genetic abnormalities were observed by Schubauer et al. [13] in the workers and patients exposed to therapeutic radiation.
Interactions between ionizing radiation and DNA produce different types of DNA lesions, including damage to nucleotide bases, DNA–DNA and DNA–protein crosslinks and alkali-labile sites, as well as SSBs and DSBs [3].
The purpose of environmental surveying and radiation dose measurement in the present study is to estimate the radiation exposure to the radiographers and visitors in the hospital.

## 2 Materials and methods

Calibrated nuclear radiation meter palm RAD 907 was used for measuring the dose rate at the place of radiographers and directly door of X-ray at general hospital in Sulaimania city-Kurdistan region-Iraq. Each of the dose rates in the tables was subtracted from the background of the places (background of place shows under the tables) and visual

inspection for radiographer's facilities were checked. We discussed with most of the radiographers and works about principle of radiation protection in their departments.

# 3 Results and discussion

Table (1) shows information of the X-ray machines. The one which used for fluoroscopy was manufactured twenty years ago while the other for routine X-ray was manufactured seven years ago. Table (2) shows general information about radiation protection tools. It is very important to estimate absorbed doses from individuals occupationally exposed to ionizing radiation in order to carry out radioprotection procedures and restrict the hazards to human health [14]. It was observed for both fluoroscopy and X-ray the radiographers were not using lead dress and paws during imaging and no one has personal monitoring badge which it is essential for protecting form X-ray. Both Deterministic and stochastic effects may result after exposure to X-ray. If radiographers work in violation of hospital safety policies and procedures they are at risk of deterministic effect. Some stochastic effects would be cancer and genetic mutation from occupational exposure to X-rays. These effects usually appear after a long time [3].
Table (3) shows information of fluoroscopy and X-ray rooms, after discussing with the workers it appears that these rooms at the beginning are not built for these purposes, so the places are poor for radiation protection.

Table 1: X-ray machine information

| Examination type | Machine name | Made in | Type of machine | Work | Tube |
|---|---|---|---|---|---|
| Fluoroscopy | Siemens | Italy (1993-1994) | Const. | Electronic | Standard |
| X-ray | Shimadzu | Japan (2006) | Const. | Electronic | Standard |

Table 2: General observation of radiation protection tools

| Examination type | Lead dress for radiographic | | Paws | | Lead dress for patient | | Glass lead | | Personal monitoring badge | | Room light |
|---|---|---|---|---|---|---|---|---|---|---|---|
| | using | Non using | using | Non using | using | Non using | exist | Not exist | using | Non-using | |
| Fluoroscopy | | * | | * | | * | * | | | * | OK |
| X-ray | | * | | * | | * | * | | | * | OK |

Table 3: X-ray rooms information

| Examination type | Room dimension/ cm | Width of wall/cm | Number of windows | Number of doors | Ventilation |
|---|---|---|---|---|---|
| Fluoroscopy | 601 x 707 | 28 | 0 | 1 | 0 |
| X-ray | 450 x 421 | 16 | 0 | 1 | 1 |

The dose rates measurement at radiographers place and in front of the direct X-ray room for routine X-ray imaging recorded in tables (4 a, b). These data obtained after subtracted from the background of the place. Most of high dose rates were in front of the direct X-ray room because the door was not lined lead completely there was some leaking under the door so the people which pass there may exposed to this dose rate. The maximum dose rate for this place was 0.339 μSv/hr which almost three times more than the background of the place. However high dose rate was not recorded for the radiographer place but the problem with the unit was the radiographers do a lot of X-ray imaging in one day and most of them have private units for X-ray imaging outside the hospital, so they receive a lot of low doses per day may lead to stochastic effect in the future.
The data for dose rate of fluoroscopy tabulated in table (5). It contains one case, only for radiographer place because the place was not safe and built inside the room which was not isolated completely. For this case maximum dose rate was 1.280 μSv/hr which is almost fourteen times more than background of the place. Diagnosis by fluoroscopy it take more time than routine X-ray, the table also shows the dose rate at the place during swallow of barium which is low but should be taken in to account.
One of another notes in this department of the hospital was two categories of personnel are not exist; medical physicist and record officer

Table (4 a): Dose rate measurement at different places

| Part of body | Position | Point of interest | Dose rate (μSv/hr) |
|---|---|---|---|
| Skull | A.P | Place for standing radiographic | 0.018 |
| | | Direct in front of door | 0.095 |
| | Lat | Place for standing radiographic | 0.026 |
| | | Direct in front of door | 0.125 |
| Shoulder | A.P | Place for standing radiographic | 0.020 |
| Cervical spine | | Direct in front of door | 0.063 |
| | Lat | Place for standing radiographic | 0.017 |
| | | Direct in front of door | 0.119 |
| Chest | P.A | Place for standing radiographic | 0.058 |
| | | Direct in front of door | 0.174 |
| | Lat | Place for standing radiographic | 0.039 |
| | | Direct in front of door | 0.293 |
| Humerus | A.P | Place for standing radiographic | 0.021 |
| | | Direct in front of door | 0.116 |
| Elbow | A.P | Place for standing radiographic | 0.012 |
| | | Direct in front of door | 0.047 |
| | Lat | Place for standing radiographic | 0.003 |
| | | Direct in front of door | 0.107 |
| Knee | A.P | Place for standing radiographic | 0.015 |
| Leg | | Direct in front of door | 0.114 |
| Foot | Lat | Place for standing radiographic | 0.014 |
| | | Direct in front of door | 0.086 |
| Hand | A.P & Lat | Place for standing radiographic | 0.026 |
| Wrist | | Direct in front of door | 0.131 |
| Femur | A.P | Place for standing radiographic | 0.032 |
| | | Direct in front of door | 0.117 |
| | Lat | Place for standing radiographic | 0.017 |
| | | Direct in front of door | 0.099 |

Table (4 b): Dose rate measurement at different places

| Part of body | Position | Point of interest | Dose rate (μSv/hr) |
|---|---|---|---|
| Forearm | A.P & Lat | Place for standing radiographic | 0.048 |
| | | Direct in front of door | 0.122 |
| Abdomen | A.P | Place for standing radiographic | 0.021 |
| | | Direct in front of door | 0.240 |
| Pelvis | A.P | Place for standing radiographic | 0.029 |
| | | Direct in front of door | 0.152 |
| Kidney Ureter Bladder (K.U.B) | A.P | Place for standing radiographic | 0.015 |
| | | Direct in front of door | 0.185 |
| Lumbar sacral spine | A.P | Place for standing radiographic | 0.030 |
| | | Direct in front of door | 0.309 |
| | Lat | Place for standing radiographic | 0.003 |
| | | Direct in front of door | 0.285 |
| Thoracic vertebrae | A.P | Place for standing radiographic | 0.026 |
| | | Direct in front of door | 0.264 |
| | Lat | Place for standing radiographic | 0.018 |
| | | Direct in front of door | 0.339 |

Background = 0.086 μSv/hr (Place for standing radiographic)
Background = 0.124 μSv/hr (Direct in front of door)

Table 5: Dose rate measurement at radiographer place for fluoroscopy

| Part of body | Position | Point of interest | Dose rate (μSv/hr) |
|---|---|---|---|
| Lumbar sacral spine | A.P | Place for standing radiographic | 1.280 |
| | Lat | Place for standing radiographic | 1.187 |
| Intravenous urography (I.V.U) | A.P | | 0.116 |
| Barium meal | | Place for standing radiographic | |
| B.enema | | | |
| B.follow through | | | |
| Barium swallow | A.P | Place for standing radiographic | 0.059 |

Background = 0.088 Sv/hr (Place for standing radiographic)

# 4 Conclusion

Radiation protection of the radiographers and the visitors are not taken into account in the hospital. Excess dose rate at the places may increase the stochastic effects in the future, so the responsible of the unit was informed about the results.

# Acknowledgements

The author would like to thank all the staffs and workers at the unit of X-ray at the hospital which helped to measure the dose rate and kindly answered the questions.

# References

[1] Radiation Protection Guidance for Hospital Staff, Prepared for Stanford Hospital and Clinics, Lucile Packard Children's Hospital and Veterans Affairs Palo Alto Health Care System, December 2010 Last revision August 23, 2012.

[2] Wall B F, Kendall G M, Edwards A A, Bouffler S, Muirhead C R, and Meara J R, "What are the risks of medical X-ray and other low dose radiation?".The British Journal of Radiology.79 (2006) 285-294.

[3] Vera Garaj-Vrhovac1 and Nevenka Kopjar "The alkaline Comet assay as biomarker in assessment of DNA damage in medical personnel occupationally exposed to ionizing radiation". Mutagenesis 18 (3) (2003) 265–271.

[4] Olisekodiaka, M.J., Bello, T.O., Onuegbu, A.J., Olowookere, C.J., Lebi, O.M., Omotayo,L.O., Igbeneghu,T., Olugbenga-Bello, A.T. "Evaluation of the Serum Total AntioxidantStatus Levels in Health Workers Exposed to Low Radiation Dose in a large Nigerian Hospital" International Journal of Research and Review in applied Science.3(2)(2009) 742-752.

[5] IAEA Biological Dosimetry – Chromosomal Aberration Analysis for Dose Assessment, International Atomic Energy Technical Report Series no. 260. IAEA, Vienna, (1986) pp. 1–69.

[6] Carrano,A.V. Chromosomal aberrations as markers of exposure and effect. J. Occup. Med., (1986) 1112–1116.

[7] Carrano,A.V. and Natarajan,A.T. Considerations for population monitoring usin cytogenetic techniques. Mutat. Res., 204 (1988) 379–406.

[8] Garaj-Vrhovac,V. and Kopjar,N. The comet assay – a new technique for detection of DNA damage in genetic toxicology studies and human biomonitoring. Period. Biol., 100 (1998) 361–366.

[9] Maluf,S.W., Passos,D.F., Bacelar,A., Speit,G. and Erdtmann,B. Assessment of DNA damage in lymphocytes of workers exposed to Xradiation using the micronucleus test and the comet assay. Environ. Mol. Mutagen., 38 (2001) 311–315.

[10] Kasuba, V., Rozgar, R.and Sentija, K., Chromosomal aberrations in medical staff occupationally exposed to X-rays: a follow-up study. Arh. Hig. Rada. Toksikol., 49 (1998) 1-8

[11] Lalic, H. and Radosevic, S.B., Chromosome aberrations in peripheral blood lymphocytes in subjects occupationally exposed to ionizing radiation or chemical clastogens. Folia. Biol., 48 (2002) 102-107.

[12] Cheriyan VD, Kurian CJ, Das B, Karuppa Swamy CV Genetic monitoring of humanpopulation from high level natural radiation areas of Kerala on the southwest coast of India. Radiat Res, 152(6) (1999) 154-158.

[13] Schubauer-Berigan MK, Wenz TB. Leukemia mortality among radiation exposed workers.Occup Med, 16(2) (2001) 271-281.

[14] Ramalho,A.T., Costa,M.L.P. and Oliveira,M.S. Conventional radiationbiologicaldosimetry using frequencies of unstable chromosome aberrations. Mutat. Res., 404 (1998) 97–100.